

## SP-NORMOPLASTIN (PT Reagent)

REF: 614 000 1 x 8 ml  
REF: 614 001 6 x 4 ml  
REF: 614 002 8 x 6 ml  
REF: 614 003 6 x 8 ml

### Intended Use

Spectrum Diagnostics SP-NORMOPLASTIN reagent is intended for prothrombin time (PT) determination.

### Background

The arrest of bleeding depends upon primary platelet plug formed along with the formation of a stable fibrin clot. Formation of this clot involves the sequential interaction of series of plasma proteins in a highly ordered and complex manner and also the interaction of these complexes with blood platelets and materials released from the tissues. Tissue thromboplastin, in the presence of calcium, is an activator which initiates the extrinsic pathway of coagulation, which includes plasma coagulation factors VII, X, V, prothrombin and fibrinogen. During oral anticoagulant therapy most of these factors are depressed, as also during the deficiencies of clotting factor activity which may be hereditary or acquired. Prothrombin time determination is the preferred method for presurgical screening, determination of congenital deficiency of factors II, VII, X, V and for monitoring of patients on oral anticoagulant therapy and as a liver function test.

### Assay Principle

Tissue thromboplastin in the presence of calcium activates the extrinsic pathway of human blood coagulation mechanism. When SP-NORMOPLASTIN reagent is added to normal anticoagulated plasma, the clotting mechanism is initiated, forming a solid gel clot within a specified period of time. The time required for clot formation would be prolonged if there is a deficiency of factors / factor activity in the extrinsic pathway of the coagulation mechanism.

### Reagent Preparation

SP-NORMOPLASTIN is a liquid ready to use Calcium Thromboplastin reagent, which is derived from rabbit brain. Each batch of reagents undergoes vigorous quality control at various stages of manufacture for its sensitivity and performance.

### Reagent Storage and Stability

Store the reagent at 2 – 8 °C (Do not freeze). The reagent is stable up to the expiry date stated on the vial label when stored capped at 2 – 8 °C. Once opened, the reagent is stable for 2 months at 2 – 8 °C, 1 week at 18 – 25 °C, 2 days at 37 °C.

### Deterioration

Failure to recover control values within the assigned range may be an indication of reagent deterioration.

### Specimen Collection and Preparation of PPP

Though no special preparation of the patient is required prior to sample collection by approved techniques, it is preferable that patients are not heavily exercised before blood collection. Fasting or only light non-fatty meals prior to blood collection provide samples with a desirable lower opacity. Withdraw blood without undue venous stasis or frothing into a plastic syringe fitted with a short needle. The venipuncture must be a "clean" one and, if there is any difficulty, take a new syringe and needle and try another vein.

Transfer the blood into anticoagulated tubes, after detaching the needle from the syringe. Do not delay mixing blood with anticoagulant. Avoid foam formation during mixing. Mix exactly nine parts of freshly collected blood with one part of trisodium citrate (0.11 mol/L, 3.2%). For occasional patients with hematocrit less than 20% or greater than 55%, this ratio must be readjusted to ensure valid result. Centrifuge immediately for 15 minutes at 1500 – 3000 rpm (approximately 1500g) on laboratory centrifuge and transfer the plasma into a clean test tube. It should be ensured that the plasma is free from platelets (PPP). Cap the test tubes to prevent deterioration of samples. Plasma must be tested preferably immediately. However if the specimens are held at 2 - 4 °C then they may be tested within 3 hours.

### SYMBOLS IN PRODUCT LABELLING

	Authorised Representative		Use by/Expiration Date
	For in-vitro diagnostic use		CAUTION. Consult instructions
	Batch Code/Lot number		for use
	Catalogue Number		Manufactured by
	Consult instructions for use		(Xi) - Irritant
	Temperature Limitation		

### Procedure

#### Manual Method

- Aspirate from the reagent vial enough reagent for immediate testing requirements in a thoroughly clean and dry test tube (Plastic test tubes are preferred).
- Bring this reagent to room temperature before prewarming at 37°C for testing purpose.
- Recap the reagent vial and replace immediately to 2 – 8°C.
- To a 12x75 mm tube add 50 µl of plasma (ppp) and place the tube in a water bath for 3 to 5 minutes at 37 °C.
- To the tube forcibly add 0.1 ml of SP-NORMOPLASTIN reagent (prewarmed at 37 °C for at least 3 minutes) and simultaneously start a stop watch. Shake the tube gently to mix contents.
- Gently tilt the tube back and forth and stop the stopwatch as soon as the first fibrin strand is visible and the gel / clot formation begins. Record the time in seconds.
- Repeat steps 4-6 for a duplicate test on the same samples.
- Find the average of the duplicate test values. This is the Prothrombin Time (PT)

### Calculation of Results

#### Manual Method

The result may be reported directly in terms of the mean of the double determination of PT of the test plasma in seconds. Or as a ratio R:

$$R = \frac{\text{Mean of the patient plasma PT in seconds}}{\text{MNPT for the reagent}^*}$$

Or as international Normalized Ratio (INR).  $INR = \left(\frac{ISI}{R}\right)^{ISI}$ , where ISI = International Sensitivity index of the reagent (Refer to reagent vial label)

\*It is recommended by WHO that MNPT (mean normal PT) should be established for each lot of PT reagents by each laboratory, since PT results are dependent on the combination of reagent lot, instrument and technique followed at each laboratory. Usually plasma from at least 20 normal healthy individuals should be used to establish the MNPT. The average of such PT results in seconds = MNPT.

### Expected Values

Normal values using SP-NORMOPLASTIN are between 10-14 seconds. Between manual and turbo densitometric instrument results a variation of 1-2 seconds may be expected.

For photo optical instruments, it is recommended that each laboratory must establish its own normal range. It is mandatory that each laboratory must establish its own MNPT for each lot of SP-NORMOPLASTIN.

Oral anticoagulant therapeutic range : INR = 2.0 – 3.5

The use of INR's enables direct comparison to be made between all results on patient plasmas regardless of interlab variations or reagent in question.

The INR is calculated as  $INR = \left(\frac{ISI}{R}\right)^{ISI}$   
Where ISI = Lot specific ISI for the reagent

$$\text{And, } R = \frac{\text{Patient PT}}{\text{Mean normal PT}}$$

Mean normal PT = Mean of the normal range that is specifically determined by each user laboratory for each lot of thromboplastin reagent with specific instrument and techniques routinely used for patient testing.

### Example:

Patient PT result = 21 seconds  
MNPT = 13.5 seconds.  
ISI of reagent = 1.5

$$R = \frac{21.0}{13.5} = 1.5$$

$$INR = (1.5)^{1.5} = 1.8$$

Alternatively the INR value can be read off directly from the SP-NORMOPLASTIN INR conversion table.

### Remarks

1. It is recommended that controls with known factor activity should be run simultaneously with each test series to validate test run.
2. Incorrect mixture of blood and tri-sodium citrate, insufficient prewarming of plasma and reagent, contaminated reagents, glassware, etc. are potential source of errors.
3. Oxalated plasma may induce prolonged clotting times.
4. Since the PT test functions correctly only at  $37 \pm 0.5$  °C, temperature of all equipments must be calibrated daily.
5. Clotting time of patients on anticoagulant therapy depends upon the type and dosage of anticoagulant and also the time lag between the specimen collected and the last dose.
6. Turbid, icteric, lipemic or grossly hemolysed samples may generate erroneous PT results.
7. Glasswares and cuvettes used in the test must be scrupulously clean and free from even traces of acids / alkalies or detergents.
8. Plasma samples held at  $4 - 8$ °C may undergo cold activation leading to a marked shortening of the PT.
9. The PT may be shortened during acute inflammatory conditions which are accompanied by increase in fibrinogen levels and also by agent such as antihistamines, butabarbital, phenobarbital, caffeine, oral contraceptives and vitamin K. the PT may be prolonged by corticosteroides, EDTA, asparaginase, clofibrate, ethanol, tetracycline, aspirin and anticoagulants such as heparin and warfarin.
10. It is important that each laboratory expresses the results in terms of INR for patients on oral anticoagulant therapy for the clinician to adjust the dosage based on INR.
11. Since the test uses platelet poor plasma, each laboratory must calibrate the necessary force and time required during centrifugation to yield the PPP. Contamination of plasma with excess platelets could falsely elevate levels of some of the factors.
12. Homogenisation of SP-NORMOPLASTIN reagent suspension before use is important to achieve accurate and consistent results.

### Precautions and Warnings

For in vitro diagnostic use only. Do not ingest or inhale. In case of contact with eyes or skin; rinse immediately with plenty of soap and water. In case of severe injuries; seek medical advice immediately.

### Waste Disposal

This product is made to be used in professional laboratories. Please consult local regulations for a correct waste disposal.

**S56:** dispose of this material and its container at hazardous or special waste collection point.

**S57:** use appropriate container to avoid environmental contamination.

**S61:** avoid release in environment. refer to special instructions/safety data sheets.

### References

1. Biggs R. and R.G. McFarlane: Human Coagulation and its disorders, Blackwell Scientific Publications, Oxford, 1962.
2. Quick A.J., Hemorrhagic disease and thrombosis, 2nd Ed., Philadelphia, Lee and Fibiger, 1966.
3. CRC Handbook Series in clinical laboratory Science, section 1: Haematology, Volume, 1980.
4. E.A. Loeliger, A.M.H.P. Van den Besselaar and S.M. Lewis, Reliability and Clinical impact of Normalization of Prothrombin Times in Oral anticoagulant control – F.K. Schattauer Verlag GmbH 1985.
5. Hull R., Hirsh H., Jay R., et al., Difference intensities of oral anticoagulant therapy in the treatment of proximal vein thrombosis, N. Engl. J. Med., 1982, 307, 1676-81.
6. W.H.O. Expert Committee on Biological standardization, No. 687, 1983.

ORDERING INFORMATION	
CATALOG NO.	QUANTITY
614 000	40 test
614 001	120 test
614 002	240 test
614 003	240 test



Spectrum For Diagnostic Industries - Free Zone

Ismailia Free Zone Industrial Area, Block 5 .

Cairo- Port said Avenue.

Ismailia, Egypt

Tel: +2 064 3488 013 - +2 064 3488 014 Fax: +2 064 3488 015

www.sdi-fz.com



MDSS GmbH

Schiffgraben 41

30175 Hannover, Germany



IFUFHR02

Rev.(2),5/5/2020