











SP-NORMOTROL

Control serum normal/borderline, for Control of Accuracy and Precision in quantitative In vitro Diagnostics

REF: 326 001 (1 vial) 1 x 5 ml
 REF: 326 002 (5 vials) 5 x 5 ml
 REF: 326 003 (6 vials) 6 x 5 ml

SYMBOLS IN PRODUCT LABELLING			
	Authorised Representative		Use by/Expiration Date
	For in-vitro diagnostic use		CAUTION. Consult instructions for use
	Batch Code/Lot number		Manufactured by
	Catalogue Number		(Xi) - Irritant
	Consult instructions for use		Temperature Limitation

Description

SP-NORMOTROL is a lyophilized human-based control serum.

Storage and stability

Unopened bottles must be stored at 2 - 8°C and are stable up to the expiration date printed on the labels

After reconstitution the serum can be used, stored at 2 - 8°C and protected from light, in between 5 days.

Bilirubin is stable 1 day, refrigerated.

The frozen serum is stable for at least 30 days at < -20°C.

Warnings and Precautions

Each individual blood donation used for this serum was found negative when tested with FDA-approved methods on HBsAg, anti-HIV 1+2 and anti-HCV. Nevertheless the serum should be treated for safety reasons always as a potentially infectious material.

Preparation

Reconstitute carefully by adding exactly 5 ml of distilled water. Allow the vial to stand for 30 minutes, swirling occasionally. Avoid foaming i.e. do not shake! For alkaline phosphatase the control serum should stand for 2 hours at +25 °C before use.

Procedure

Refer to the package inserts of the reagent kits

Assay Values and Ranges

Assay values for analytes for which approved reference methods are available were determined according to Guidelines of the German Federal Medical Council [Bundesärztekammer] from 1987 (reference method values) [3].

Ranges of acceptance were calculated as assigned value ± three times the maximum tolerable deviation of a single value according to the Guidelines of the German Federal Medical Council [4].

Expected Values

1. The expected value of specific assays are provided on the assay value sheet accompanying each kit, and are lot specific.
2. The expected values are obtained using replicate assay of each manufactured lot of SP-Normotrol.
3. The individual laboratory values should fall within the expected values.
4. It must however be noted that each laboratory should establish its own normal values and reference range according to GLP.

Literature

1. Röhle G, Siekmann L. Quality assurance of quantitative determination. In: Thomas L, editor. Clinical laboratory diagnostics. 1st ed. Frankfurt:TH-Books Verlagsgesellschaft; 1998. p. 1393-1401
2. Biosafety in Microbiological and Biomedical Laboratories. U.S. Department of Health and Human Services, Washington 1993 (HHS Publication No. [CDC] 93-8395).
3. Richtlinien der Bundesärztekammer zur Qualitätssicherung in medizinischen Laboratorien. Deutsches Ärzteblatt 1988;85: B519-B532.
4. Richtlinie der Bundesärztekammer zur Qualitätssicherung quantitativer laboratoriumsmedizinischer Untersuchungen. Deutsches Ärzteblatt 2002;98:A 2747-59.

 Spectrum For Diagnostic Industries - Free Zone
 Ismailia Free Zone Industrial Area, Block 5 .
 Cairo- Port said Avenue.
 Ismailia, Egypt
 Tel: +2 064 3488 013 - +2 064 3488 014 Fax: +2 064 3488 015
 www.sdi-fz.com



MDSS GmbH
 Schiffgraben 41
 30175 Hannover, Germany

