

Triglycerides – Liquizyme GPO-PAP (Single Reagent)

REF: 314 001	(2 x 25 ml)	50 test
REF: 314 002	(4 x 25 ml)	100 test
REF: 314 003	(4 x 30 ml)	120 test
REF: 314 004	(10 x 15 ml)	150 test
REF: 314 005	(4 x 50 ml)	200 test
REF: 314 006	(4 x 60 ml)	240 test
REF: 314 007	(4 x100 ml)	400 test
REF: 314 008	(5 x100 ml)	500 test
REF: 314 009	(8 x100 ml)	800 test
REF: 314 010	(4 x250 ml)	1000 test

Intended Use

Spectrum Diagnostics liquizyme Triglycerides reagent is intended for the in-vitro quantitative, diagnostic determination of triglycerides in human serum on both automated and manual systems.

Background

Triglycerides are the main lipids present in the human plasma; the others are the cholesterol, phospholipids and nonesterified fatty acids. They are formed in the intestinal mucosa by the esterification of glycerol and fatty acids. Triglycerides measurements are used in the diagnosis and treatment of patients with diabetes mellitus, liver obstruction, nephrosis and other diseases involving lipid metabolism. The measurement of serum triglycerides is important in the diagnosis of hyperlipoproteinemia and in the prediction, detection and monitoring of atherosclerosis.

Method

GPO-PAP-enzymatic colorimetric method.

Assay Principle

The series of the reaction involved in the assay system is as follows:

- Triglycerides are hemolyzed by lipoprotein lipase (LPL) to glycerol



- Glycerol is then phosphorylated to glycerol-3-phosphate by ATP in a reaction catalyzed by glycerol kinase (GK).



- The oxidation of glycerol-3-phosphate is catalyzed by glycerol phosphate oxidase (GPO) to form dihydroxyacetone phosphate and hydrogen peroxide (H₂O₂).



- In the presence of peroxidase (POD), hydrogen peroxide effects the oxidative coupling of 4-chlorophenol and 4-aminoantipyrine (4AAP) to form a red color quinoneimine dye which is measured at 546 nm.



Reagents

Standard Triglycerides (ST)

200 mg/dL 2.29 mmol/L

Reagent (R)

Pipes Buffer pH 7.0	50 mmol/L
4-chlorophenol	6.0 mmol/L
Magnesium aspartate	>0.5 mmol/L
Lipase	>10 K U/L
Peroxidase	>2.0 KU/L
4-Aminoantipyrine	1.0 mmol/L
Glycerol-3-phosphate oxidase	>3.5 K U/L
Glycerol kinase	>750 U/L
ATP	1.0 mmol/L
Sodium Azide	8.0 mmol/L

For further information, refer to the Triglycerides reagent material safety data sheet..

SYMBOLS IN PRODUCT LABELLING			
	Authorised Representative		Use by/Expiration Date
	For in-vitro diagnostic use		CAUTION. Consult instructions for use
	Batch Code/Lot number		Manufactured by
	Catalogue Number		(Xi) - Irritant
	Consult instructions for use		Temperature Limitation

Precautions and Warnings

Do not ingest or inhale. In case of contact with eyes or skin; rinse immediately with plenty of soap and water. In case of severe injuries; seek medical advice immediately.

Reagent (R) contains sodium azide which may react with copper or lead plumbing.

Reagent Preparation, Storage and Stability

Spectrum triglycerides reagents are supplied ready-to-use and stable up to the expiry date labeled on the bottles and glycerol free collection device. Recommended anticoagulants are EDTA or heparin at levels of 1mg and 0.2 mg/dl whole blood, respectively.

Deterioration

The reagent is normally clear or pale pink. Do not use liquizyme triglyceride reagent if it is turbid or if the absorbance is greater than 0.2 at 546 nm.

Specimen Collection and Preservation

Patients should be fasting for 10 to 14 hours before blood is drawn. Samples must be drawn in a soap and glycerol free collection device. Recommended anticoagulants are EDTA or heparin at levels of 1mg and 0.2 mg/dl whole blood, respectively.

Triglycerides in serum samples remain stable for 7 days at 4 °C, for 3 months at -20 °C, and for years at -70 °C.

System Parameters

Wavelength	Hg 546 nm (500 – 550 nm)
Optical path	1 cm
Assay type	End-point
Direction	Increase
Sample : Reagent Ratio	1 : 100
e.g.: Reagent volume	1 ml
Sample volume	10 µl
Temperature	15 – 25 °C or 37 °C
Zero adjustment	Reagent blank
Incubation time	10 minutes at 15 – 25 °C or 5 minutes at 37 °C
Reagent Blank Limits	Low 0.00 AU High 0.2 AU
Sensitivity	5 mg/dL (0.057 mmol/L)
Linearity	1000 mg/dL (11.45 mmol/L)

Procedure

	Blank	Standard	Sample
Reagent	1.0 ml	1.0 ml	1.0 ml
Standard	-----	10 µl	-----
Sample	-----	-----	10 µl

Mix and incubate for 5 minutes at 37 °C or 10 minutes at 15 – 25°C. Measure absorbance of specimen (A_{specimen}) and standard (A_{standard}) against reagent blank within 30 minutes.

Calculation

$$\text{Serum Triglycerides conc. (mg/dL)} = \frac{A_{\text{specimen}}}{A_{\text{standard}}} \times 200$$

Quality Control

Normal and abnormal commercial control serum of known concentrations should be analyzed with each run.

Performance Characteristics

Precision

Within run (Repeatability)

	Level 1	Level 2
n	20	20
Mean (mg/dL)	155.1	245.8
SD	2.03	1.85
CV%	1.31	0.75

Run to run (Reproducibility)

	Level 1	Level 2
n	20	20
Mean (mg/dL)	156	246.5
SD	2.2	1.9
CV%	1.4	0.87

Methods Comparison

A comparison between Spectrum Diagnostics Triglycerides reagent and a commercial reagent of the same methodology was performed on 20 human sera. A correlation of 0.967 was obtained.

Sensitivity

When run as recommended, the minimum detection limit of the assay is 5 mg/dL (0.057 mmol/L).

Linearity

The reaction is linear up to triglycerides concentration of 1000 mg/dL; specimens showing higher concentration should be diluted 1+1 using physiological saline and repeat the assay (result × 2).

Interfering Substances

Haemolysis

No significant interference up to a haemoglobin level of 6.0 g/L (0.36 mmol/L).

Icterus

Bilirubin levels higher than 171 µmol/L (10 mg/dL) decrease the apparent Triglycerides concentration significantly.

Drugs

Of the drugs tested in-vitro, methyl dopa and levodopa cause artificially low Triglycerides values at the tested drug Level.

Others

Physiological ascorbic acid concentration doesn't interfere with the test. Ascorbic acid levels higher than 114 µmol/l (2 mg/dL) decrease the apparent Triglycerides concentration significantly.

Expected Values

Females	35 -135 mg/dL	(0.4 – 1.54 mmol/L)
Males	40 -160 mg/dL	(0.45 – 1.82 mmol/L)

The following limits are recommended for the recognition of the risk factor hypertriglyceridemia:

Suspicious	above 150 mg/dL	(1.71 mmol/L)
Elevated	above 200 mg/dL	(2.28 mmol/L)

Spectrum Diagnostics does not interpret the results of a clinical laboratory procedure; interpretation of the results is considered the responsibility of qualified medical personnel. All indications of clinical significance are supported by literature references.

Dynamic Range

5 - 1000 mg/dL (0.057 - 11.45 mmol/L)

Waste Disposal

This product is made to be used in professional laboratories. Please consult local regulations for a correct waste disposal.

S56: dispose of this material and its container at hazardous or special waste collection point.

S57: use appropriate container to avoid environmental contamination.

S61: avoid release in environment. refer to special instructions/safety data sheets.

References

- Bucolo G, David H : Quantitative determination of serum triglycerides by the use of the enzymes. Clin Chem 19 : 475, 1973
- Chowdhury RF, Rodman H, Bleicher SJ : Glycerol like contamination of commercial blood sampling tubes. J Clin Pathol 12: 116, 1971
- MGowan MW, Artiss JD, Standbergh DR, Zak B. A peroxidase-coupled method for colorimetric determination of serum triglycerides. Clin Chem ;29:538-452 ;1983.
- Stein EA; Lipids , lipoproteins, and apolipoproteins. In : NW Tietz , ed. Fundamentals of clinical chemistry, 3 rd ed. Philadelphia : WB Saunders; 448 ; 1987.
- Tietz NW, Boden T, Stepleton JD : An improved method for the determination of lipase in serum. Am J Clin Pathol 31: 148, 1959
- Young DS et al, Clin Chem. 21 ; 1975

ORDERING INFORMATION

CATALOG NO.	QUANTITY
314 001	2 x 25 ml
314 002	4 x 25 ml
314 003	4 x 30 ml
314 004	10 x 15 ml
314 005	4 x 50 ml
314 006	4 x 60 ml
314 007	4 x100 ml
314 008	5 x100 ml
314 009	8 x100 ml
314 010	4 x250 ml



Spectrum For Diagnostic Industries - Free Zone

Ismailia Free Zone Industrial Area, Block 5 .

Cairo- Port said Avenue.

Ismailia, Egypt

Tel: +2 064 3488 013 - +2 064 3488 014 Fax: +2 064 3488 015

www.sdi-fz.com



MDSS GmbH
Schiffgraben 41
30175 Hannover, Germany



IFUFCC39

Rev.(2), 21/7/2020